



## CERTIFICATE OF ANALYSIS

**Product: Karl Fischer Reagent 5**  
Single component reagent pyridine free 1 ml  $\cong$  5 mg H<sub>2</sub>O  
**Order No.:** 246.05.01000.01

**Batch No/Lot No:** 246.05.09-H31.21  
**Quantity:** 1 Lit  
**Date of Production:** 31/08/2021  
**Retest Date:** 31/08/2023

Row	Test Item	Specification	Results	Reference
1	Efficiency	$\geq 5$ mg H <sub>2</sub> O/ml at the time of manufacture	<b>5.25</b>	Merck Standard

**QC. Manager:** Ms. Iravani

**Status:**

QC.PC.WIN.105/COA.03



**Qualified Person:** Dr. Mojallali

**Date:** 04.09.2021

