

Vis Array Spectrometer



- Simultaneous obtained of all spectra
- Integration time controlling using software
- Mathematical processing (Averaging, Sum, Smooth ...)
- Time Monitoring of Spectra

Specifications of Array Spectrometer

Optical Specifications

Wavelength range	380-800 nm (380-1100)
Straylight	< 0.03 %
Resolution	< + 1.0 nm
Wavelength Accuracy	< ± 0.1 nm
Wavelength Reproducibility	< ± 0.02 nm for 10 consecutive scans (NIST 2034)
Photometric Accuracy	< ± 0.05 A at 440.0 465.0 590.0 635.0 nm(NIST 930e)
Photometric noise	<0.05 A 80 scan at 0 A, 500 nm
Photometric stability	<0.01 A/min at 0 A, 500 nm
Baseline flatness	<0.01A 0.5 second blank, 0.5 second scan at 0 A, 500 nm
Scan time	2-2000 ms
Typical Scan Time	0.1 s
Detector	1000 Multichannel

Physical Specifications

Communication	USB 2
Software	Windows XP , 7, 8
Height × width × depth	22 × 12 × 27.5 cm
Weight	4 kg
Power	220V AC 50/60 HZ

Tell: +98 21 88357692

Fax: +98 21 88357803

Address: No.115, Dr. Fathi shaghghi St, Yosef abad, Tehran, Iran

Teifsanje Tajhiz Pishrafteh Co. Ltd.

Email: info@teifsanje.ir

Website: www.teifsanje.ir

