

بسمه تعالی

شرایط کالیبراسیون :

موارد	صحت	واحد
Voltage measurement relative error	5-200 \pm 5	uV %
Sensitivity adjustment relative error	5-100 \pm 5	uV %
Time Interval relative error	0.05- 5 \pm 5	S - S%
Record Rate relative error	15-60 \pm 5	uS
Residual Record error	0.5-10 \pm 5	uV/S %
Time marker & Internal Calibrator Relative error	2-1000 \pm 2	Us - S %
Overshoot	+10	uV%
Time Constant error	\pm 10	S %
Frequency Response Amplitude	0.16-75 \pm 10	Hz%
Input impedance	>20 \pm 10	M Ω %
Common mode rejection Ratio	200-10000 \pm 5	%
Base Line With	50-1500 \pm 10	uV%
Base Line Drift	100 \pm 10	uV%