

## Nanodrop UV-Vis Array Spectrophotometer



- Nano & Micro volume analysis
- Simultaneous obtained of all spectra
- No evaporation of Nanodrop
- High-precision optics
- Standard Wavelength calibration of spectrophotometer



## Specifications of Nanodrop Array Spectrophotometer

Optical Specifications	
Wavelength range	200-800 nm (190-1100 nm)
Straylight	< 0.03 %
Resolution	< + 1.0 nm
Wavelength Accuracy	< ± 0.1 nm
Wavelength Reproducibility	< ± 0.02 nm for 10 consecutive scans (NIST 2034)
Photometric Accuracy	< ± 0.01 A at 235.0 257.0 313.0 nm
Photometric noise	<0.01 A 30 scan at 0 A, 500 nm
Photometric stability	<0.01 A/min at 0 A, 500 nm
Baseline flatness	<0.01A 0.5 second blank, 0.5 second scan at 0 A, 500 nm
Scan time	2-1000 ms
Typical Scan Time	0.1 s
Light Source	Deuterium + Halogen /Tungsten lamp
Detector	2000 Multichannel
Sample Specifications	
Sample volume	700 - 4000 nL or 3000-4000 nL
Light path	0.2mm or 1mm
Physical Specifications	
Communication	USB 2
Software	Windows XP , 7, 8
Height × width × depth	13 × 33 × 50 cm
Weight	10 kg
Power	220V AC 50/60 HZ

Tell : (021)88357692

Fax : (021)88357803

Address : No.115, Dr. Fathi shaghghi St, Yosef abad, Tehran, Iran

Email : [info@spectrometer.ir](mailto:info@spectrometer.ir)

Website : [www.spectrometer.ir](http://www.spectrometer.ir)

